



# Steinbeis

- Transferzentrum  
an der Hochschule Mannheim


## Certificate

The GAMMA-SCOUT® unit No. **071106** was used to measure the environmental radiation for three days. The results were compared with those of a reference device and found to deviate from these by less than  $\pm 5\%$ .

The reference devices were measured at the Insitut für Physikalische Chemie und Radiochemie at the Hochschule Mannheim (public university for applied technology) in a dose equivalent range up to 1000  $\mu\text{Sv/h}$ , and determined suitable.

The reference devices are regularly checked.

The measuring performance of the GAMMA-SCOUT® unit No. **071106** can therefore be certified as being very good.



Prof. Dr. E. Fohr



# Steinbeis

- Transferzentrum  
an der Hochschule Mannheim


## Certificate

The GAMMA-SCOUT® unit No. **071243** was used to measure the environmental radiation for three days. The results were compared with those of a reference device and found to deviate from these by less than  $\pm 5\%$ .

The reference devices were measured at the Insitut für Physikalische Chemie und Radiochemie at the Hochschule Mannheim (public university for applied technology) in a dose equivalent range up to 1000  $\mu\text{Sv/h}$ , and determined suitable.

The reference devices are regularly checked.

The measuring performance of the GAMMA-SCOUT® unit No. **071243** can therefore be certified as being very good.



Prof. Dr. E. Fohr