


Outline

- 1 Introduction
- 2 Tokamak GOLEM - engineering scheme
- 3 Tokamak GOLEM - basic diagnostics
- 4 Tokamak GOLEM - operation
- 5 Data handling @ the Tokamak GOLEM
- 6 Closings
- 7 Appendix

Basic characteristics



- Major radius $R_0 = 0.4 \text{ m}$
- Minor radius $r_0 = 0.1 \text{ m}$
- Plasma radius $a = 0.085 \text{ m}$
- Toroidal magnetic field $B_t < 0.5 \text{ T}$
- Plasma current $I_p < 8 \text{ kA}$
- Plasma density
 $n \approx 0.2 - 3 \times 10^{19} / \text{m}^{-3}$
- Electron temperature $T_e < 100 \text{ eV}$
- Ion temperature $T_i < 50 \text{ eV}$
- Length of the discharge $\tau < 20 \text{ ms}$

Tokamak GOLEM for education - historical background

Kurchatov Institute near Moscow,
Soviet Union
1960: **TM1-MH**



1974


Institute of Plasma Physics
Czech republic

CASTOR

COMPASS


2008

Czech Technical University Prague
Czech republic
GOLEM



Culham Centre for Fusion Energy
Great Britain
1989: **COMPASS-D**

2006





GOLEM



The new location of the tokamak is just next to the old Prague Jewish cemetery where Rabi Loew (Golem builder) is buried, and that is why it was renamed GOLEM (and also for the symbol of potential power you get if you know the magic). Interestingly, here in Prague, where the Golem legend originated, Golem is not perceived as a symbol of evil, but rather as a symbol of power which might be useful but is very challenging to handle. To learn more of the Golem legend, see e.g. [wikipedia](https://en.wikipedia.org/wiki/Golem).

The global schematic overview of the tokamak GOLEM experiment

LEVEL 1



Infrastructure room (below tokamak) 10/16



Infrastructure room (below tokamak) 10/16

Current drive CD field
and toroidal magnetic Bt field
circuits

To the tokamak
GOLEM

Rotary
pump

Vacuum
control

Current drive CD
capacitors

Plasma
stabilization


power
supply
2kV

Toroidal
magnetic field B
capacitors


fire
protection
system




Tokamak room (North) 10/16




Tokamak room (North) 10/16



Tokamak room (South) 10/16




Tokamak room (South) 10/16



Outline

- 1 Introduction
- 2 Tokamak GOLEM - engineering scheme
- 3 Tokamak GOLEM - basic diagnostics
- 4 Tokamak GOLEM - operation
- 5 Data handling @ the Tokamak GOLEM
- 6 Closings
- 7 Appendix

Plasma in Tokamak (GOLEM) - the least to do




To do:

- Evacuate the chamber
- Fill in the working gas
- Preionization
- Toroidal magnetic field to confine plasma
- Toroidal electric field to breakdown neutral gas into plasma
- Toroidal electric field to heat the plasma
- Plasma positioning
- Diagnostics

Plasma in Tokamak (GOLEM) - the least to do

LEVEL 1




To do:

- **Evacuate the chamber**
- Fill in the working gas
- Preionization
- Toroidal magnetic field to confine plasma
- Toroidal electric field to breakdown neutral gas into plasma
- Toroidal electric field to heat the plasma
- Plasma positioning
- Diagnostics

Plasma in Tokamak (GOLEM) - the least to do

LEVEL 1




To do:

- Evacuate the chamber
- **Fill in the working gas**
- Preionization
- Toroidal magnetic field to confine plasma
- Toroidal electric field to breakdown neutral gas into plasma
- Toroidal electric field to heat the plasma
- Plasma positioning
- Diagnostics

Plasma in Tokamak (GOLEM) - the least to do

LEVEL 1




To do:

- Evacuate the chamber
- Fill in the working gas
- **Preionization**
- Toroidal magnetic field to confine plasma
- Toroidal electric field to breakdown neutral gas into plasma
- Toroidal electric field to heat the plasma
- Plasma positioning
- Diagnostics

Plasma in Tokamak (GOLEM) - the least to do

LEVEL 1




To do:

- Evacuate the chamber
- Fill in the working gas
- Preionization
- **Toroidal magnetic field to confine plasma**
- Toroidal electric field to breakdown neutral gas into plasma
- Toroidal electric field to heat the plasma
- Plasma positioning
- Diagnostics

Plasma in Tokamak (GOLEM) - the least to do

LEVEL 1




To do:

- Evacuate the chamber
- Fill in the working gas
- Preionization
- Toroidal magnetic field to confine plasma
- **Toroidal electric field to breakdown neutral gas into plasma**
- **Toroidal electric field to heat the plasma**
- Plasma positioning
- Diagnostics




Outline

- 1 Introduction
- 2 Tokamak GOLEM - engineering scheme
- 3 Tokamak GOLEM - basic diagnostics**
- 4 Tokamak GOLEM - operation
- 5 Data handling @ the Tokamak GOLEM
- 6 Closings
- 7 Appendix



Tokamak GOLEM - basic diagnostics






Loop voltage U_l




Toroidal magnetic field B_t




Total current I_{ch+p}



Basic diagnostics @ tokamak GOLEM



Plasma x vacuum discharge



Outline

- 1 Introduction
- 2 Tokamak GOLEM - engineering scheme
- 3 Tokamak GOLEM - basic diagnostics
- 4 Tokamak GOLEM - operation**
- 5 Data handling @ the Tokamak GOLEM
- 6 Closings
- 7 Appendix


Remote operation

*Tokamak Golem **REMOTE** for PROMO (Level I)*

The smallest & oldest operational tokamak with the biggest control room in the world



LEVEL 1



Shot homepage

GOLEM » Shot #22471 »



Diagnostics

- ✓ Interferometer
- ✓ Spectrometer
- ✗ FastCamera
- ✓ HXR

Analysis

- ✓ Shot-Homepage

DAS

- ✓ TektronixDPO
- ✓ Nistandard
- ✓ Papouch_St
- ✓ Papouch_Ko
- ✓ Nloctopus

Vacuum log

Other

- Data
- References
- About
- Wiki
- Utilities

Navigation

- Next
- Previous
- Current

Tokamak GOLEM - Shot Database - 22471


Date: 2016-09-29 - 14:33:57
Session: TrainingCourses/Universities/Uni_Belgrade.rs/2016/
Comment: Standard discharge

Basic parameters: (compare)

- Gas pressure p_{ch} : 0.42->20.39 mPa (request: 20 mPa) [WIKI](#)
- Working gas: H
- Preionization: Upper el. gun
- Chamber temperature: 27.20 C
- C_B capacitors charged to: 800 V, triggered 5.0 ms [WIKI](#)
- C_{BD} capacitors charged to: 0 V, triggered 5.0 ms [WIKI](#)
- C_{CD} capacitors charged to: 400 V, triggered 6.0 ms [WIKI](#)
- C_{ST} capacitors charged to: 0 V, triggered 5.0 ms [WIKI](#)
- Probability of breakdown: 85% [WIKI](#)
- Time since session beginning: 0:07:50 h

Plasma parameters:

- Plasma life time 14.8 [ms] (from 7.8 to 22.6)
- Mean toroidal magnetic field B_t : 0.23 T [WIKI](#)
- Mean plasma current: 3.60 kA [WIKI](#)
- Mean Uloop: 5.92 V [WIKI](#)
- Break down voltage: 9.6 V [WIKI](#)
- Ohmic heating power: 21.33 kW
- Q edge: 2.9 [WIKI](#)
- Electron temperature: 41.1 eV [WIKI](#)
- Line electron density: 5.52 [$10^{17} \cdot m^{-2}$] [WIKI](#)



Outline

- 1 Introduction
- 2 Tokamak GOLEM - engineering scheme
- 3 Tokamak GOLEM - basic diagnostics
- 4 Tokamak GOLEM - operation
- 5 Data handling @ the Tokamak GOLEM**
- 6 Closings
- 7 Appendix

GOLEM basic Data Acquisition System (DAS)


- $U_I, U_{B_t}, U_{I_{p+ch}}, I_{rad}$
- $\Delta t = 1\mu s / f = 1MHz$.
- Integration time = 40 ms, thus DAS produces 6 columns x 40000 rows data file.
- Discharge is triggered at 5th millisecond after DAS to have a zero status identification.

Data file example, DAS $\Delta t = 1\mu s / f = 1MHz$ (neutral gas into plasma breakdown focused)


t	$\approx U_I$	$\approx U_{\frac{dB_T}{dt}}$	$\approx U_{\frac{d(I_{p+ch})}{dt}}$	$\approx I_{rad}$
:	:	:	:	:
:	:	:	:	:
first	$\approx$	7405	lines ..	
:	:	:	:	:
:	:	:	:	:
0.007383	1.53931	0.390015	0.048828	0.001831
0.007384	1.53686	0.395508	0.067749	0.00061
0.007385	1.54053	0.391235	0.079956	0.00061
0.007386	1.53686	0.38147	0.072632	0
0.007387	1.54297	0.397949	0.059204	0.00061
0.007388	1.54053	0.384521	0.05249	0.00061
0.007389	1.54053	0.39856	0.068359	0.001221
0.00739	1.54053	0.393677	0.082397	0.001221
0.007391	1.53809	0.38208	0.072632	0.001221
0.007392	1.54297	0.400391	0.056763	0.00061
0.007393	1.54419	0.383911	0.053101	0.00061
0.007394	1.53931	0.397339	0.068359	0.001221
0.007395	1.54297	0.391846	0.084229	0.00061
0.007396	1.54541	0.394897	0.074463	0.00061
0.007397	1.54297	0.388184	0.056763	0.001221
0.007398	1.54297	0.391846	0.056763	0.00061
0.007399	1.54297	0.394287	0.06897	0.00061
:	:	:	:	:
:	:	:	:	:
next	$\approx$	32500	lines ..	
:	:	:	:	:
:	:	:	:	:




Plot 4665 U_l graph




python




gnuplot




mathematica



idl



octave



matlab

Data access

All the recorded data and the settings for each shot are available at the GOLEM website. The root directory for the files is:

```
http://golem.fjfi.cvut.cz/shots/<#ShotNo>/
```

Actually last discharge has the web page:

```
http://golem.fjfi.cvut.cz/shots/0.
```

Particular data from DAS or specific diagnostics have the format:

```
http://golem.fjfi.cvut.cz/utls/data/<#ShotNo>/<identifier>.
```

GNU Wget

GNU Wget is a free software package for retrieving files using HTTP, HTTPS and FTP, the most widely-used Internet protocols. It is a non-interactive commandline tool, so it may easily be called from scripts, cron jobs, terminals without X-Windows support, etc.

- Runs on most UNIX-like operating systems as well as Microsoft Windows.
- Homepage: <http://www.gnu.org/software/wget/>
- Basic usage:
 - To get U_l : `wget http://golem.fjfi.cvut.cz/utis/data/<#ShotNo>/loop_voltage`
 - To get whole shot: `wget -r -nH -cut-dirs=3 -no-parent -l2 -Pshot http://golem.fjfi.cvut.cz/shots/<#ShotNo>`

Matlab

```
ShotNo=22471;
baseURL='http://golem.fjfi.cvut.cz/utis/data/';
identifier='loop_voltage';
%Create a path to data
dataURL=strcat(baseURL,int2str(ShotNo),'/',identifier);
% Write data from GOLEM server to a local file
urlwrite(dataURL,identifier);
% Load data
data = load(identifier, '\t');
% Plot and save the graph
plot(data(:,1)*1000, data(:,2), '.');
xlabel('Time [ms]')
ylabel('UL [V]')
saveas(gcf, 'plot', 'jpeg');
exit;
```

Jupyter (python)


```
import matplotlib.pyplot as plt
from numpy import loadtxt
from urllib import urlopen

baseURL = "http://golem.fjfi.cvut.cz/Utils/data/"
ShotNo = 22471
identifier = "loop_voltage"
#Create a path to data
dataURL = urlopen(baseURL+ str(ShotNo) + '/' + identifier)
#Load data from GOLEM server
data=loadtxt(dataURL, delimiter='\t')
#Plot the graph
plt.plot(data[:,0], data[:,1], 'k-')
plt.savefig('graph.jpg')
plt.show()
```

Gnuplot

```
set macros;  
ShotNo = "22471";  
baseURL = "http://golem.fjfi.cvut.cz/utils/data/";  
identifier = "loop_voltage";  
#Create a path to data  
DataURL= "@baseURL@ShotNo/@identifier";  
#Write data from GOLEM server to a local file  
!wget -q @DataURL;  
#Plot the graph from a local file  
set datafile separator "\t";  
plotstyle = "with_lines_linestyle_-1"  
plot 'loop_voltage' using 1:2 @plotstyle;  
exit;  
  
# command line execution:  
# gnuplot Uloop.gp -persist
```


Excel



File→Open→

<http://golem.fjfi.cvut.cz/utis/data/<#ShotNo>/<identifier>>

Spreadsheets (Excel and others)


are not recommended, only tolerated.

Outline

- 1 Introduction
- 2 Tokamak GOLEM - engineering scheme
- 3 Tokamak GOLEM - basic diagnostics
- 4 Tokamak GOLEM - operation
- 5 Data handling @ the Tokamak GOLEM
- 6 Closings**
- 7 Appendix

- Everything via <http://golem.fjfi.cvut.cz/ASPNF>
 - Session: Thursday, 2 Feb
 - This presentation.
 - Control rooms (login: golem, password: XXXX).
 - Contact: Vojtech Svoboda, +420 737673903.
 - possible chat: vojtech.svob@gmail.com
 - skype: tokamak.golem.
 - essential: Test data processing in advance (e.g. #22471)

Tokamak GOLEM - virtual model



Acknowledgement

The financial support by FUSENET@EU, Ministry of education@CzechRep, SGS15/164/OHK4/2T/14@CzechTechUniv, CRP F1.30.14@IAEA, RVO68407700@NuclearFaculty, .

Special thanks to the GOLEM team (students, teachers, technicians)

Ondrej Grover, Jaroslav Krbec, Jindrich Kocman, Michal Odstrcil, Tomas Odstrcil, Gergo Pokol, Jan Stockel, Gabriel Vondrasek, and others ...

Winter school of Plasma Physics - Marianska 2016 (Toroidal field coil 4 ITER, cooling test)



Outline

- 1 Introduction
- 2 Tokamak GOLEM - engineering scheme
- 3 Tokamak GOLEM - basic diagnostics
- 4 Tokamak GOLEM - operation
- 5 Data handling @ the Tokamak GOLEM
- 6 Closings
- 7 Appendix**

References I



V. Svoboda, B. Huang, J. Mlynar, G.I. Pokol, J. Stockel, and G Vondrasek.

Multi-mode Remote Participation on the GOLEM Tokamak.
Fusion Engineering and Design, 86(6-8):1310–1314, 2011.



Brotankova, J.

Study of high temperature plasma in tokamak-like experimental devices.

PhD. thesis 2009.



Tokamak GOLEM team.

Tokamak GOLEM at the Czech Technical University in Prague.
<http://golem.fjfi.cvut.cz>, 2007.

References II



J. Wesson.

Tokamaks, volume 118 of *International Series of Monographs on Physics*.

Oxford University Press Inc., New York, Third Edition, 2004.

Physical Quantities @ the tokamak GOLEM

Loop Voltage: U_l [V]

Total (plasma+chamber) current: I_{p+ch} [A]

Chamber current: I_{ch} [A]

Plasma current: I_p [A]

Plasma resistivity: R_p [Ω]

Plasma heating power: P_{OH} [W]

Total plasma energy content: W_p [J]

Chamber resistivity: R_{ch} [Ω]

Electron temperature: T_e [eV]

Energy confinement time: τ_E [s]

Plasma volume: $V_p = 0,057$ [m^3]

Rogowski coil calibration constant: $K_{RogCoil} = 5.3 \cdot 10^6$ [A/Vs]

Loop Voltage calibration constant: $K_{LoopVoltage} = 5.5$ [-]

Boltzmann constant : $k_B = 1.38064852 \cdot 10^{-23}$ [J/K]